

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122021\
 Data File : VN070239.D
 Acq On : 20 Dec 2021 14:37
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 21 04:01:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1260283	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2406811	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2874861	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1363042	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	971704	49.94	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.88%
35) Dibromofluoromethane	8.021	113	684793	45.42	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.84%
50) Toluene-d8	10.440	98	3218424	53.17	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.34%
62) 4-Bromofluorobenzene	12.731	95	1525590	69.57	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	139.14%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122021\
Data File : VN070239.D
Acq On : 20 Dec 2021 14:37
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 21 04:01:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

