

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122021\
 Data File : VN070249.D
 Acq On : 20 Dec 2021 18:52
 Operator : JC/MD
 Sample : M5044-06
 Misc : 5.82g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DA-6-4.5

Quant Time: Dec 21 04:06:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1129091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2072243	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2339193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1298428	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	879777	50.47	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.94%			
35) Dibromofluoromethane	8.018	113	611072	47.08	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 94.16%			
50) Toluene-d8	10.440	98	2949313	56.59	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 113.18%			
62) 4-Bromofluorobenzene	12.731	95	1480525	78.41	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 156.82%#			
Target Compounds					Qvalue	
39) Methylcyclohexane	9.453	83	3064639	130.60	ug/l #	80
73) Isopropylbenzene	12.578	105	1348148	11.67	ug/l	98
78) n-propylbenzene	12.918	91	3658009	28.46	ug/l	95
85) sec-Butylbenzene	13.500	105	1146996	10.27	ug/l #	61
89) n-Butylbenzene	13.940	91	4083432	57.67	ug/l	88
95) Naphthalene	15.504	128	5326473	81.24	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122021\
Data File : VN070249.D
Acq On : 20 Dec 2021 18:52
Operator : JC/MD
Sample : M5044-06
Misc : 5.82g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DA-6-4.5

Quant Time: Dec 21 04:06:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

