

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122021\
 Data File : VN070254.D
 Acq On : 20 Dec 2021 20:59
 Operator : JC/MD
 Sample : M5129-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 48-CI-1

Quant Time: Dec 21 04:09:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1266317	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2390558	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2840242	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1235483	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1029706	52.67	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.34%
35) Dibromofluoromethane	8.021	113	712253	47.57	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.14%
50) Toluene-d8	10.440	98	3346710	55.66	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.32%
62) 4-Bromofluorobenzene	12.731	95	1507482	69.21	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	138.42%#
Target Compounds						
						Qvalue
16) Acetone	4.291	43	158861	13.51	ug/l	98
20) Methylene Chloride	5.095	84	79123	5.06	ug/l	90
43) Isopropyl Acetate	8.563	43	134456	2.68	ug/l #	87
95) Naphthalene	15.507	128	26088	4.38	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122021\
Data File : VN070254.D
Acq On : 20 Dec 2021 20:59
Operator : JC/MD
Sample : M5129-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
48-CI-1

Quant Time: Dec 21 04:09:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

