

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122022\
 Data File : VN075888.D
 Acq On : 20 Dec 2022 10:48
 Operator : JC\MD
 Sample : VN1220MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1220MBL01

Quant Time: Dec 21 01:26:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.230	168	376388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	578887	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	491384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	213474	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	185840	45.385	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.760%
35) Dibromofluoromethane	8.177	113	170274	47.497	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.000%
50) Toluene-d8	10.571	98	648250	48.363	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.720%
62) 4-Bromofluorobenzene	12.853	95	207683	46.781	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122022\
 Data File : VN075888.D
 Acq On : 20 Dec 2022 10:48
 Operator : JC\MD
 Sample : VN1220MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1220MBL01

Quant Time: Dec 21 01:26:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

