

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122022\
 Data File : VN075898.D
 Acq On : 20 Dec 2022 14:59
 Operator : JC\MD
 Sample : N6148-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMPOSITE-3052

Quant Time: Dec 21 01:28:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	387596	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	608269	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	528935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	215795	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	196641	46.634	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.260%
35) Dibromofluoromethane	8.177	113	177378	47.088	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.180%
50) Toluene-d8	10.571	98	705467	50.090	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.180%
62) 4-Bromofluorobenzene	12.853	95	222831	47.769	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.540%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	16018	11.020	ug/l	97
20) Methylene Chloride	5.283	84	22320	5.166	ug/l	93
37) Ethyl Acetate	7.571	43	31253	7.776	ug/l	99
43) Isopropyl Acetate	8.694	43	209685	31.625	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122022\
Data File : VN075898.D
Acq On : 20 Dec 2022 14:59
Operator : JC\MD
Sample : N6148-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMPOSITE-3052

Quant Time: Dec 21 01:28:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

