

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122022\
 Data File : VN075900.D
 Acq On : 20 Dec 2022 15:47
 Operator : JC\MD
 Sample : N6150-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-K

Quant Time: Dec 21 01:28:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	400045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	636845	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	556323	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	228331	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	203561	46.773	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.540%
35) Dibromofluoromethane	8.171	113	190134	48.210	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.420%
50) Toluene-d8	10.571	98	738427	50.077	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.160%
62) 4-Bromofluorobenzene	12.853	95	229800	47.052	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.100%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	15079	10.051	ug/l	92
20) Methylene Chloride	5.289	84	11436	2.238	ug/l	94
37) Ethyl Acetate	7.571	43	31938	7.589	ug/l	98
43) Isopropyl Acetate	8.694	43	265023	38.178	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122022\
Data File : VN075900.D
Acq On : 20 Dec 2022 15:47
Operator : JC\MD
Sample : N6150-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-K

Quant Time: Dec 21 01:28:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

