

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122023\
 Data File : VN080538.D
 Acq On : 20 Dec 2023 11:31
 Operator : JC\MD
 Sample : VN1220MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1220MBL01

Quant Time: Dec 21 02:38:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	198879	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	378111	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	378982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	171725	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	144497	56.885	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.780%
35) Dibromofluoromethane	8.165	113	130088	54.263	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.520%
50) Toluene-d8	10.565	98	439951	49.206	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.420%
62) 4-Bromofluorobenzene	12.847	95	178230	52.496	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122023\
Data File : VN080538.D
Acq On : 20 Dec 2023 11:31
Operator : JC\MD
Sample : VN1220MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1220MBL01

Quant Time: Dec 21 02:38:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

