

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122023\
 Data File : VN080549.D
 Acq On : 20 Dec 2023 16:07
 Operator : JC\MD
 Sample : 05874-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 21 02:44:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	222701	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	406446	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	446581	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	196657	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	166265	58.453	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.900%
35) Dibromofluoromethane	8.165	113	142669	55.362	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.720%
50) Toluene-d8	10.565	98	497611	51.775	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.560%
62) 4-Bromofluorobenzene	12.847	95	205277	56.247	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122023\
Data File : VN080549.D
Acq On : 20 Dec 2023 16:07
Operator : JC\MD
Sample : 05874-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 21 02:44:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

