

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122023\
 Data File : VN080550.D
 Acq On : 20 Dec 2023 16:31
 Operator : JC\MD
 Sample : 05873-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 215-0209(TB-07)RE

Quant Time: Dec 21 02:44:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	228381	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	432421	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	453120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	203397	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	167193	57.318	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	114.640%	
35) Dibromofluoromethane	8.165	113	140622	51.290	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.580%	
50) Toluene-d8	10.571	98	501353	49.031	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.060%	
62) 4-Bromofluorobenzene	12.847	95	200054	51.523	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	103.040%	
Target Compounds						
16) Acetone	4.424	43	6908	13.356	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122023\
Data File : VN080550.D
Acq On : 20 Dec 2023 16:31
Operator : JC\MD
Sample : 05873-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
215-0209(TB-07)RE

Quant Time: Dec 21 02:44:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

