

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122023\
 Data File : VN080553.D
 Acq On : 20 Dec 2023 17:41
 Operator : JC\MD
 Sample : 05942-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-121923

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/21/2023
 Supervised By : Mahesh Dadoda 12/21/2023

Quant Time: Dec 21 02:48:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	197235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	368078	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	396688	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	173171	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	149995	59.542	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 119.080%			
35) Dibromofluoromethane	8.165	113	127235	54.519	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.040%			
50) Toluene-d8	10.565	98	446761	51.330	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.660%			
62) 4-Bromofluorobenzene	12.853	95	178366	53.968	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 107.940%			
Target Compounds						
16) Acetone	4.424	43	3683m	8.245	ug/l	
18) Methyl Acetate	5.018	43	3133	1.090	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122023\
Data File : VN080553.D
Acq On : 20 Dec 2023 17:41
Operator : JC\MD
Sample : 05942-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-121923

Quant Time: Dec 21 02:48:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/21/2023
Supervised By :Mahesh Dadoda 12/21/2023

