

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122023\
 Data File : VN080555.D
 Acq On : 20 Dec 2023 18:29
 Operator : JC\MD
 Sample : 05862-10 20X
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-SOLIDS-20231213-01

Quant Time: Dec 21 02:49:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	213902	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	405858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	436210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	203822	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	158966	58.186	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 116.380%	
35) Dibromofluoromethane	8.165	113	134342	52.206	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.420%	
50) Toluene-d8	10.565	98	495897	51.672	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.340%	
62) 4-Bromofluorobenzene	12.853	95	202087	55.453	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.900%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	1706	3.522	ug/l	89
67) Ethyl Benzene	11.965	91	21151	1.185	ug/l #	83
68) m/p-Xylenes	12.071	106	7745	1.175	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122023\
Data File : VN080555.D
Acq On : 20 Dec 2023 18:29
Operator : JC\MD
Sample : 05862-10 20X
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-SOLIDS-20231213-01

Quant Time: Dec 21 02:49:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

