

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122121\
 Data File : VN070259.D
 Acq On : 21 Dec 2021 11:28
 Operator : JC/MD
 Sample : VN1221WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1221WBL01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/22/2021
 Supervised By : Semsettin Yesilyurt 12/23/2021

Quant Time: Dec 21 15:23:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	993501	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1641388	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1418178	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	486406	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	846833	55.21	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 110.42%			
35) Dibromofluoromethane	8.021	113	562875	54.75	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 109.50%			
50) Toluene-d8	10.441	98	2137047	51.77	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.54%			
62) 4-Bromofluorobenzene	12.731	95	691292	46.22	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 92.44%			
Target Compounds						Qvalue
20) Methylene Chloride	5.103	84	7781m	0.63	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122121\
Data File : VN070259.D
Acq On : 21 Dec 2021 11:28
Operator : JC/MD
Sample : VN1221WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1221WBL01

Quant Time: Dec 21 15:23:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/22/2021
Supervised By :Semsettin Yesilyurt 12/23/2021

