

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122121\
 Data File : VN070276.D
 Acq On : 21 Dec 2021 18:50
 Operator : JC/MD
 Sample : M5182-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Dec 21 20:12:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	781508	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1337128	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1148955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	372210	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	681977	56.518	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 113.040%	
35) Dibromofluoromethane	8.021	113	445624	53.206	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.420%	
50) Toluene-d8	10.440	98	1690475	50.266	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.540%	
62) 4-Bromofluorobenzene	12.731	95	531930	43.660	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 87.320%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122121\
Data File : VN070276.D
Acq On : 21 Dec 2021 18:50
Operator : JC/MD
Sample : M5182-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Dec 21 20:12:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

