

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122121\
 Data File : VN070277.D
 Acq On : 21 Dec 2021 19:16
 Operator : JC/MD
 Sample : M5182-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EQUIPMENT-BLANK

Quant Time: Dec 21 20:12:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 12/22/2021
 Supervised By :Semsettin Yesilyurt 12/23/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	794589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1345849	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1165651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	370072	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	687647	56.050	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 112.100%			
35) Dibromofluoromethane	8.024	113	456061	54.099	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 108.200%			
50) Toluene-d8	10.440	98	1718349	50.764	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.520%			
62) 4-Bromofluorobenzene	12.731	95	531169	43.315	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 86.620%			
Target Compounds					Qvalue	
16) Acetone	4.293	43	67002	9.079	ug/l	100
20) Methylene Chloride	5.098	84	8339m	0.850	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122121\
Data File : VN070277.D
Acq On : 21 Dec 2021 19:16
Operator : JC/MD
Sample : M5182-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EQUIPMENT-BLANK

Quant Time: Dec 21 20:12:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/22/2021
Supervised By :Semsettin Yesilyurt 12/23/2021

