

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122121\
 Data File : VN070278.D
 Acq On : 21 Dec 2021 19:41
 Operator : JC/MD
 Sample : PB141540TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141540TB

Quant Time: Dec 21 20:12:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	670455	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1226878	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1037314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	347802	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	674927	65.199	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	130.400%
35) Dibromofluoromethane	8.021	113	429419	55.878	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	111.760%
50) Toluene-d8	10.440	98	1350038	43.751	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	87.500%
62) 4-Bromofluorobenzene	12.731	95	441551	39.498	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	79.000%
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	127401	20.459	ug/l	97
18) Methyl Acetate	4.867	43	22748	1.210	ug/l #	51
20) Methylene Chloride	5.095	84	77432	9.354	ug/l	87
25) 2-Butanone	7.348	43	28545	3.374	ug/l #	85
43) Isopropyl Acetate	8.563	43	149874	5.826	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122121\
Data File : VN070278.D
Acq On : 21 Dec 2021 19:41
Operator : JC/MD
Sample : PB141540TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141540TB

Quant Time: Dec 21 20:12:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

