



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122121\
Data File : VN070256.D
Acq On : 21 Dec 2021 9:45
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Title : SW846 8260
Last Update : Fri Dec 03 08:46:47 2021



AutoFind: Scans 4004, 4005, 4006; Background Corrected with Scan 3986

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	68373	PASS
75	95	30	60	54.3	165248	PASS
95	95	100	100	100.0	304299	PASS
96	95	5	9	6.5	19899	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.2	228736	PASS
175	174	5	9	7.7	17605	PASS
176	174	95	101	96.3	220203	PASS
177	176	5	9	6.7	14644	PASS