

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122122\
 Data File : VN075910.D
 Acq On : 21 Dec 2022 11:20
 Operator : JC\MD
 Sample : VN1221WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1221WBL01

Quant Time: Dec 21 14:15:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	390725	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	617149	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	532526	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	224314	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	197252	46.404	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.800%
35) Dibromofluoromethane	8.171	113	179286	46.910	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.820%
50) Toluene-d8	10.571	98	696115	48.714	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.420%
62) 4-Bromofluorobenzene	12.853	95	221298	46.757	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122122\
Data File : VN075910.D
Acq On : 21 Dec 2022 11:20
Operator : JC\MD
Sample : VN1221WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1221WBL01

Quant Time: Dec 21 14:15:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

