

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122122\
 Data File : VN075919.D
 Acq On : 21 Dec 2022 15:05
 Operator : JC\MD
 Sample : N6164-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWOW-BR-04-147-166-122022

Quant Time: Dec 22 03:02:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	357113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	555755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	487064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	196994	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	180871	46.556	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.120%
35) Dibromofluoromethane	8.177	113	164865	47.902	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.800%
50) Toluene-d8	10.571	98	635719	49.402	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.800%
62) 4-Bromofluorobenzene	12.853	95	196692	46.149	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.300%
Target Compounds						
52) Toluene	10.635	92	1152388	125.437	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122122\
Data File : VN075919.D
Acq On : 21 Dec 2022 15:05
Operator : JC\MD
Sample : N6164-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWOW-BR-04-147-166-122022

Quant Time: Dec 22 03:02:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

