

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122122\
 Data File : VN075920.D
 Acq On : 21 Dec 2022 15:29
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 22 03:02:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	378023	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	593788	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	509099	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	212563	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	188919	45.937	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.880%
35) Dibromofluoromethane	8.177	113	172925	47.026	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.060%
50) Toluene-d8	10.571	98	681992	49.604	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.200%
62) 4-Bromofluorobenzene	12.853	95	211383	46.420	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.840%
Target Compounds						
52) Toluene	10.635	92	3365	0.343	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122122\
Data File : VN075920.D
Acq On : 21 Dec 2022 15:29
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 22 03:02:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

