

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
 Data File : VN080565.D
 Acq On : 21 Dec 2023 11:43
 Operator : JC\MD
 Sample : VN1221MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1221MBL01

Quant Time: Dec 22 03:56:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	232475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	436219	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	452706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	194995	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160065	53.908	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.820%
35) Dibromofluoromethane	8.165	113	140637	50.849	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.700%
50) Toluene-d8	10.565	98	507037	49.155	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.320%
62) 4-Bromofluorobenzene	12.847	95	196418	50.147	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
Data File : VN080565.D
Acq On : 21 Dec 2023 11:43
Operator : JC\MD
Sample : VN1221MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1221MBL01

Quant Time: Dec 22 03:56:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

