

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
 Data File : VN080566.D
 Acq On : 21 Dec 2023 12:07
 Operator : JC\MD
 Sample : VN1221WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1221WBL01

Quant Time: Dec 22 03:57:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	259880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	479336	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	516130	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	221056	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	182038	54.843	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.680%
35) Dibromofluoromethane	8.165	113	162545	53.483	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.960%
50) Toluene-d8	10.565	98	583550	51.484	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.960%
62) 4-Bromofluorobenzene	12.847	95	226574	52.642	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
Data File : VN080566.D
Acq On : 21 Dec 2023 12:07
Operator : JC\MD
Sample : VN1221WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1221WBL01

Quant Time: Dec 22 03:57:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

