

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
 Data File : VN080572.D
 Acq On : 21 Dec 2023 14:40
 Operator : JC\MD
 Sample : 05896-02DL 20X
 Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 525DL

Quant Time: Dec 22 04:00:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	266290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	510849	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	547451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	256611	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	198006	58.218	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 116.440%	
35) Dibromofluoromethane	8.171	113	174690	53.934	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.860%	
50) Toluene-d8	10.570	98	643895	53.304	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.600%	
62) 4-Bromofluorobenzene	12.847	95	273914	59.715	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 119.440%	
Target Compounds						Qvalue
16) Acetone	4.424	43	134498	223.026	ug/l	99
52) Toluene	10.629	92	111921	11.585	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
Data File : VN080572.D
Acq On : 21 Dec 2023 14:40
Operator : JC\MD
Sample : 05896-02DL 20X
Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
525DL

Quant Time: Dec 22 04:00:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

