

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
 Data File : VN080573.D
 Acq On : 21 Dec 2023 15:04
 Operator : JC\MD
 Sample : 05896-01DL 40X
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1104DL

Quant Time: Dec 22 04:00:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	280308	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	516296	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	551140	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	270601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	195636	54.644	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.280%			
35) Dibromofluoromethane	8.165	113	174220	53.221	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.440%			
50) Toluene-d8	10.565	98	639749	52.402	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.800%			
62) 4-Bromofluorobenzene	12.847	95	274160	59.138	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 118.280%			
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	12502	19.694	ug/l	99
39) Methylcyclohexane	9.606	83	6405	1.152	ug/l	92
52) Toluene	10.629	92	303525	31.086	ug/l	99
67) Ethyl Benzene	11.965	91	29663	1.316	ug/l	93
68) m/p-Xylenes	12.065	106	40278	4.835	ug/l	92
69) o-Xylene	12.394	106	11935	1.465	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
Data File : VN080573.D
Acq On : 21 Dec 2023 15:04
Operator : JC\MD
Sample : 05896-01DL 40X
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1104DL

Quant Time: Dec 22 04:00:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

