

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
 Data File : VN080576.D
 Acq On : 21 Dec 2023 16:15
 Operator : JC\MD
 Sample : 05914-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-12000

Quant Time: Dec 22 04:02:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	263909	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	509097	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	528401	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	225847	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	177232	52.580	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 105.160%		
35) Dibromofluoromethane	8.165	113	161382	49.996	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.000%		
50) Toluene-d8	10.565	98	605549	50.302	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.600%		
62) 4-Bromofluorobenzene	12.853	95	239036	52.291	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 104.580%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	38846	64.996	ug/l	88
20) Methylene Chloride	5.277	84	15628	4.886	ug/l #	89
43) Isopropyl Acetate	8.688	43	186563	31.395	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
Data File : VN080576.D
Acq On : 21 Dec 2023 16:15
Operator : JC\MD
Sample : 05914-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-12000

Quant Time: Dec 22 04:02:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

