

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
 Data File : VN080577.D
 Acq On : 21 Dec 2023 16:38
 Operator : JC\MD
 Sample : 05926-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MR-319-0-2

Quant Time: Dec 22 04:03:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	313078	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	642603	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	802408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	332909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	210501	52.642	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 105.280%		
35) Dibromofluoromethane	8.165	113	196016	48.110	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.220%		
50) Toluene-d8	10.571	98	911129	59.961	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 119.920%#		
62) 4-Bromofluorobenzene	12.847	95	332422	57.612	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 115.220%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	44353	62.555	ug/l	99
20) Methylene Chloride	5.277	84	17229	4.541	ug/l #	88
43) Isopropyl Acetate	8.688	43	230133	30.681	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
Data File : VN080577.D
Acq On : 21 Dec 2023 16:38
Operator : JC\MD
Sample : 05926-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MR-319-0-2

Quant Time: Dec 22 04:03:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

