

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
 Data File : VN080580.D
 Acq On : 21 Dec 2023 17:49
 Operator : JC\MD
 Sample : 05975-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT-GWDUP01-122023

Quant Time: Dec 22 04:05:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	375536	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	789804	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	760254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	317359	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	276433	57.633	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 115.260%			
35) Dibromofluoromethane	8.165	113	254639	50.850	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.700%			
50) Toluene-d8	10.565	98	894163	47.878	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.760%			
62) 4-Bromofluorobenzene	12.847	95	346810	48.903	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.800%			
Target Compounds						
16) Acetone	4.441	43	6150	7.231	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
Data File : VN080580.D
Acq On : 21 Dec 2023 17:49
Operator : JC\MD
Sample : 05975-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT-GWDUP01-122023

Quant Time: Dec 22 04:05:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

