

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
 Data File : VN080582.D
 Acq On : 21 Dec 2023 18:37
 Operator : JC\MD
 Sample : 05975-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DS01-GW-122023

Quant Time: Dec 22 04:05:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	347237	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	711268	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	549337	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	266398	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	247360	55.774	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.540%
35) Dibromofluoromethane	8.165	113	233615	51.803	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.600%
50) Toluene-d8	10.565	98	610374	36.291	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	72.580%#
62) 4-Bromofluorobenzene	12.853	95	235742	36.912	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	73.820%#
Target Compounds						Qvalue
16) Acetone	4.430	43	32026	40.726	ug/l	99
52) Toluene	10.630	92	77484	5.760	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
Data File : VN080582.D
Acq On : 21 Dec 2023 18:37
Operator : JC\MD
Sample : 05975-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DS01-GW-122023

Quant Time: Dec 22 04:05:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

