

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
 Data File : VN080590.D
 Acq On : 21 Dec 2023 21:46
 Operator : JC\MD
 Sample : 05926-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MR-319-0-2

Quant Time: Dec 22 04:10:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	325403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	652148	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	688885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	302033	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	222161	53.454	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.900%
35) Dibromofluoromethane	8.171	113	205819	49.776	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.560%
50) Toluene-d8	10.571	98	775731	50.304	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.600%
62) 4-Bromofluorobenzene	12.847	95	305520	52.174	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.340%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	51872	70.389	ug/l	99
18) Methyl Acetate	5.024	43	8095	1.707	ug/l	91
20) Methylene Chloride	5.283	84	18542	4.702	ug/l	92
43) Isopropyl Acetate	8.688	43	243820	32.030	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
Data File : VN080590.D
Acq On : 21 Dec 2023 21:46
Operator : JC\MD
Sample : 05926-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MR-319-0-2

Quant Time: Dec 22 04:10:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

