

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
 Data File : VN080566.D
 Acq On : 21 Dec 2023 12:07
 Operator : JC\MD
 Sample : VN1221WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1221WBL01

Quant Time: Dec 22 03:57:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	259880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	479336	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	516130	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	221056	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	182038	54.843	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.680%
35) Dibromofluoromethane	8.165	113	162545	53.483	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.960%
50) Toluene-d8	10.565	98	583550	51.484	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.960%
62) 4-Bromofluorobenzene	12.847	95	226574	52.642	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.280%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122123\
Data File : VN080566.D
Acq On : 21 Dec 2023 12:07
Operator : JC\MD
Sample : VN1221WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1221WBL01

Quant Time: Dec 22 03:57:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN080566.D
Acq: 21 Dec 2023 12:07




Instrument :
MSVOA_N
ClientSampleId :
VN1221WBL01


Tgt Ion:168 Resp: 259880
Ion Ratio Lower Upper
168 100
99 65.9 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 54.843 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN080566.D
Acq: 21 Dec 2023 12:07

Tgt Ion: 65 Resp: 182038
Ion Ratio Lower Upper
65 100
67 50.4 0.0 106.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN080566.D

Acq: 21 Dec 2023 12:07

Instrument :

MSVOA_N

ClientSampleId :

VN1221WBL01





Tgt Ion:114 Resp: 479336

Ion Ratio Lower Upper

114 100

63 20.6 0.0 36.4

88 18.2 0.0 30.6



#35

Dibromofluoromethane

Concen: 53.483 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN080566.D

Acq: 21 Dec 2023 12:07




Tgt Ion:113 Resp: 162545


Ion Ratio Lower Upper

113 100

111 103.4 82.6 123.8

192 18.9 12.7 19.1









#50
Toluene-d8
Concen: 51.484 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN080566.D
Acq: 21 Dec 2023 12:07




Instrument :
MSVOA_N
ClientSampleId :
VN1221WBL01


Tgt Ion: 98 Resp: 583550
Ion Ratio Lower Upper
98 100
100 66.2 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 52.642 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN080566.D
Acq: 21 Dec 2023 12:07

Tgt Ion: 95 Resp: 226574
Ion Ratio Lower Upper
95 100
174 75.1 0.0 143.8
176 71.8 0.0 135.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN080566.D
Acq: 21 Dec 2023 12:07

Instrument :
MSVOA_N
ClientSampleId :
VN1221WBL01

Tgt Ion:117 Resp: 516130
Ion Ratio Lower Upper
117 100
82 59.1 44.0 66.0
119 33.7 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN080566.D
Acq: 21 Dec 2023 12:07

Tgt Ion:152 Resp: 221056
Ion Ratio Lower Upper
152 100
115 64.9 32.3 96.8
150 156.4 0.0 355.8

