

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN122218\
 Data File : VN053139.D
 Acq On : 22 Dec 2018 18:05
 Operator : MD\SY
 Sample : J6377-14ME
 Misc : 4.66g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 P001-SD005-0612-01ME

Quant Time: Dec 24 09:27:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1155714	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1758699	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1597249	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	666963	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	549921	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	
35) Dibromofluoromethane	7.59	113	437810	54.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.00%	
50) Toluene-d8	10.09	98	2166069	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
62) 4-Bromofluorobenzene	12.40	95	725486	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
Target Compounds						
68) m/p-Xylenes	11.62	106	369157	15.948	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN122218\
Data File : VN053139.D
Acq On : 22 Dec 2018 18:05
Operator : MD\SY
Sample : J6377-14ME
Misc : 4.66g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
P001-SD005-0612-01ME

Quant Time: Dec 24 09:27:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

