

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122220\
 Data File : VN065220.D
 Acq On : 23 Dec 2020 1:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 23 03:48:23 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	50.000	51.822	-3.6	94	0.00
3 P	Chloromethane	50.000	52.886	-5.8	98	0.00
4 C	Vinyl Chloride	50.000	53.231	-6.5#	97	0.00
5 T	Bromomethane	50.000	51.526	-3.1	97	0.00
6 T	Chloroethane	50.000	52.476	-5.0	96	0.00
7 T	Trichlorofluoromethane	50.000	51.213	-2.4	95	0.00
8 T	Diethyl Ether	50.000	50.403	-0.8	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.139	-2.3	97	0.00
10 T	Methyl Iodide	50.000	53.623	-7.2	98	0.00
11 T	Tert butyl alcohol	250.000	259.683	-3.9	97	0.00
12 CM	1,1-Dichloroethene	50.000	49.535	0.9#	96	0.00
13 T	Acrolein	250.000	228.606	8.6	86	0.00
14 T	Allyl chloride	50.000	50.737	-1.5	95	0.00
15 T	Acrylonitrile	250.000	266.160	-6.5	98	0.00
16 T	Acetone	250.000	223.763	10.5	84	0.00
17 T	Carbon Disulfide	50.000	52.675	-5.3	97	0.00
18 T	Methyl Acetate	50.000	52.041	-4.1	97	0.00
19 T	Methyl tert-butyl Ether	50.000	51.745	-3.5	98	0.00
20 T	Methylene Chloride	50.000	48.238	3.5	99	0.00
21 T	trans-1,2-Dichloroethene	50.000	52.213	-4.4	98	0.00
22 T	Diisopropyl ether	50.000	51.620	-3.2	99	0.00
23 T	Vinyl Acetate	250.000	264.266	-5.7	97	0.00
24 P	1,1-Dichloroethane	50.000	53.075	-6.2	99	0.00
25 T	2-Butanone	250.000	252.296	-0.9	93	0.00
26 T	2,2-Dichloropropane	50.000	42.766	14.5	82	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.508	-5.0	99	0.00
28 T	Bromochloromethane	50.000	49.054	1.9	100	0.00
29 T	Tetrahydrofuran	250.000	257.641	-3.1	98	0.00
30 C	Chloroform	50.000	52.477	-5.0#	99	0.00
31 T	Cyclohexane	50.000	48.731	2.5	96	0.00
32 T	1,1,1-Trichloroethane	50.000	51.944	-3.9	98	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.014	-6.0	101	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	99	0.00
35 S	Dibromofluoromethane	50.000	51.726	-3.5	101	0.00
36 T	1,1-Dichloropropene	50.000	49.462	1.1	97	0.00
37 T	Ethyl Acetate	50.000	49.500	1.0	95	0.00
38 T	Carbon Tetrachloride	50.000	50.150	-0.3	98	0.00
39 T	Methylcyclohexane	50.000	49.107	1.8	94	0.00
40 TM	Benzene	50.000	50.351	-0.7	99	0.00
41 T	Methacrylonitrile	50.000	45.937	8.1	94	0.00
42 TM	1,2-Dichloroethane	50.000	50.032	-0.1	99	0.00
43 T	Isopropyl Acetate	50.000	50.448	-0.9	97	0.00
44 TM	Trichloroethene	50.000	49.731	0.5	97	0.00
45 C	1,2-Dichloropropane	50.000	50.376	-0.8#	98	0.00
46 T	Dibromomethane	50.000	50.430	-0.9	99	0.00
47 T	Bromodichloromethane	50.000	50.448	-0.9	98	0.00
48 T	Methyl methacrylate	50.000	51.813	-3.6	101	0.00
49 T	1,4-Dioxane	1000.000	972.207	2.8	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122220\
 Data File : VN065220.D
 Acq On : 23 Dec 2020 1:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 23 03:48:23 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	51.316	-2.6	101	0.00
51 T	4-Methyl-2-Pentanone	250.000	254.989	-2.0	98	0.00
52 CM	Toluene	50.000	51.061	-2.1#	99	0.00
53 T	t-1,3-Dichloropropene	50.000	48.489	3.0	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.491	3.0	96	0.00
55 T	1,1,2-Trichloroethane	50.000	50.474	-0.9	98	0.00
56 T	Ethyl methacrylate	50.000	50.123	-0.2	97	0.00
57 T	1,3-Dichloropropane	50.000	50.806	-1.6	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	246.261	1.5	97	0.00
59 T	2-Hexanone	250.000	251.669	-0.7	95	0.00
60 T	Dibromochloromethane	50.000	50.773	-1.5	99	0.00
61 T	1,2-Dibromoethane	50.000	50.536	-1.1	98	0.00
62 S	4-Bromofluorobenzene	50.000	48.026	3.9	100	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	48.699	2.6	99	0.00
65 PM	Chlorobenzene	50.000	49.827	0.3	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.527	-1.1	98	0.00
67 C	Ethyl Benzene	50.000	49.807	0.4#	98	0.00
68 T	m/p-Xylenes	100.000	99.781	0.2	98	0.00
69 T	o-Xylene	50.000	50.152	-0.3	98	0.00
70 T	Styrene	50.000	50.564	-1.1	97	0.00
71 P	Bromoform	50.000	51.254	-2.5	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	47.176	5.6	98	0.00
74 T	N-aryl acetate	50.000	49.682	0.6	97	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.176	3.6	99	0.00
76 T	1,2,3-Trichloropropane	50.000	44.276	11.4	95	0.00
77 T	Bromobenzene	50.000	48.471	3.1	99	0.00
78 T	n-propylbenzene	50.000	46.847	6.3	96	0.00
79 T	2-Chlorotoluene	50.000	47.054	5.9	95	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.083	5.8	97	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.006	10.0	89	0.00
82 T	4-Chlorotoluene	50.000	47.197	5.6	96	0.00
83 T	tert-Butylbenzene	50.000	47.985	4.0	98	0.00
84 T	1,2,4-Trimethylbenzene	50.000	47.782	4.4	97	0.00
85 T	sec-Butylbenzene	50.000	47.232	5.5	97	0.00
86 T	p-Isopropyltoluene	50.000	47.421	5.2	97	0.00
87 T	1,3-Dichlorobenzene	50.000	49.239	1.5	99	0.00
88 T	1,4-Dichlorobenzene	50.000	48.348	3.3	97	0.00
89 T	n-Butylbenzene	50.000	47.241	5.5	94	0.00
90 T	Hexachloroethane	50.000	49.843	0.3	98	0.00
91 T	1,2-Dichlorobenzene	50.000	50.044	-0.1	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.950	-1.9	98	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.882	2.2	102	0.00
94 T	Hexachlorobutadiene	50.000	44.677	10.6	95	0.00
95 T	Naphthalene	50.000	49.144	1.7	106	0.00
96 T	1,2,3-Trichlorobenzene	50.000	48.567	2.9	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122220\
Data File : VN065220.D
Acq On : 23 Dec 2020 1:07
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 23 03:48:23 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220w.M
Quant Title : Sw846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6