

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122221\
 Data File : VN070291.D
 Acq On : 22 Dec 2021 16:01
 Operator : JC/MD
 Sample : M5184-07RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-3RE

Quant Time: Dec 22 16:40:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	982289	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1640561	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	1444118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	485201	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	985696	64.991	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 129.980%	
35) Dibromofluoromethane	8.021	113	664752	64.689	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 129.380%	
50) Toluene-d8	10.443	98	2483427	60.187	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 120.380%	
62) 4-Bromofluorobenzene	12.731	95	800708	53.565	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.140%	
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	55200	6.050	ug/l	99
25) 2-Butanone	7.345	43	175012	14.118	ug/l	92
29) Tetrahydrofuran	7.694	42	98944	13.145	ug/l	88
52) Toluene	10.507	92	8896	0.305	ug/l	95
59) 2-Hexanone	11.089	43	19054	1.129	ug/l	89
84) 1,2,4-Trimethylbenzene	13.369	105	8421	0.254	ug/l	91
89) n-Butylbenzene	13.946	91	10276	0.388	ug/l	97
93) 1,2,4-Trichlorobenzene	15.265	180	7902	3.813	ug/l	90
95) Naphthalene	15.507	128	36287	5.393	ug/l	99
96) 1,2,3-Trichlorobenzene	15.702	180	9012	3.541	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122221\
Data File : VN070291.D
Acq On : 22 Dec 2021 16:01
Operator : JC/MD
Sample : M5184-07RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-3RE

Quant Time: Dec 22 16:40:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

