

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122221\
 Data File : VN070294.D
 Acq On : 22 Dec 2021 17:18
 Operator : JC/MD
 Sample : M5182-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EQUIPMENT-BLANKRE

Quant Time: Dec 23 04:37:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	820679	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	1422792	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1232681	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	404201	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	784514	61.913	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	123.820%
35) Dibromofluoromethane	8.021	113	522946	58.678	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	117.360%
50) Toluene-d8	10.443	98	1958553	54.731	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.460%
62) 4-Bromofluorobenzene	12.734	95	607999	46.899	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.800%
Target Compounds						
						Qvalue
16) Acetone	4.291	43	69749	9.151	ug/l	95
20) Methylene Chloride	5.090	84	8707	0.859	ug/l	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122221\
Data File : VN070294.D
Acq On : 22 Dec 2021 17:18
Operator : JC/MD
Sample : M5182-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EQUIPMENT-BLANKRE

Quant Time: Dec 23 04:37:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

