

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122222\
 Data File : VN075937.D
 Acq On : 22 Dec 2022 11:53
 Operator : JC\MD
 Sample : VN1222WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1222WBL01

Quant Time: Dec 23 03:58:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	367941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	583414	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	511636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	215116	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	190509	47.593	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.180%
35) Dibromofluoromethane	8.177	113	173920	48.137	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.280%
50) Toluene-d8	10.571	98	672015	49.747	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.500%
62) 4-Bromofluorobenzene	12.853	95	213657	47.753	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122222\
Data File : VN075937.D
Acq On : 22 Dec 2022 11:53
Operator : JC\MD
Sample : VN1222WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1222WBL01

Quant Time: Dec 23 03:58:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

