

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN122222\  
 Data File : VN075940.D  
 Acq On : 22 Dec 2022 13:15  
 Operator : JC\MD  
 Sample : N6182-03  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 7 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_N  
**ClientSampleId :**  
 TB-01-122122

Quant Time: Dec 23 04:01:12 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N121422W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Dec 14 13:03:41 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.230	168	342200	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	536588	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	466503	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	190699	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	175662	47.185	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.380%
35) Dibromofluoromethane	8.171	113	160588	48.326	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.660%
50) Toluene-d8	10.571	98	617597	49.709	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.420%
62) 4-Bromofluorobenzene	12.853	95	192434	46.763	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN122222\  
 Data File : VN075940.D  
 Acq On : 22 Dec 2022 13:15  
 Operator : JC\MD  
 Sample : N6182-03  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 TB-01-122122

Quant Time: Dec 23 04:01:12 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N121422W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Dec 14 13:03:41 2022  
 Response via : Initial Calibration

