

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN12222\
 Data File : VN075941.D
 Acq On : 22 Dec 2022 13:39
 Operator : JC\MD
 Sample : N6182-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWOW-BR-04-185-206-122122

Quant Time: Dec 23 04:01:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	364252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	570256	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	490742	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	200758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	189664	47.862	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.720%
35) Dibromofluoromethane	8.171	113	172225	48.768	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.540%
50) Toluene-d8	10.570	98	649487	49.189	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.380%
62) 4-Bromofluorobenzene	12.853	95	203790	46.599	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.200%
Target Compounds						
52) Toluene	10.635	92	34268	3.635	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN12222\
Data File : VN075941.D
Acq On : 22 Dec 2022 13:39
Operator : JC\MD
Sample : N6182-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWOW-BR-04-185-206-122122

Quant Time: Dec 23 04:01:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

