

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122222\
 Data File : VN075942.D
 Acq On : 22 Dec 2022 14:03
 Operator : JC\MD
 Sample : N6182-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWOW-BR-04-185-206-122122-FD

Quant Time: Dec 23 04:02:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	356514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	545194	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	466781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	187355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	181555	46.810	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.620%
35) Dibromofluoromethane	8.177	113	163322	48.373	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.740%
50) Toluene-d8	10.571	98	638541	50.583	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.160%
62) 4-Bromofluorobenzene	12.853	95	192009	45.923	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.840%
Target Compounds						
52) Toluene	10.635	92	36177	4.014	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN12222\
Data File : VN075942.D
Acq On : 22 Dec 2022 14:03
Operator : JC\MD
Sample : N6182-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWOW-BR-04-185-206-122122-FD

Quant Time: Dec 23 04:02:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

