

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122222\
 Data File : VN075946.D
 Acq On : 22 Dec 2022 15:39
 Operator : JC\MD
 Sample : N6168-06 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1691

Quant Time: Dec 23 04:03:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	328111	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	515033	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	443571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	183432	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	172233	48.251	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.500%
35) Dibromofluoromethane	8.177	113	155909	48.882	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.760%
50) Toluene-d8	10.571	98	597290	50.086	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.180%
62) 4-Bromofluorobenzene	12.853	95	183356	46.422	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122222\
Data File : VN075946.D
Acq On : 22 Dec 2022 15:39
Operator : JC\MD
Sample : N6168-06 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1691

Quant Time: Dec 23 04:03:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

