

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122222\
 Data File : VN075947.D
 Acq On : 22 Dec 2022 16:03
 Operator : JC\MD
 Sample : N6168-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1839

Quant Time: Dec 23 04:03:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	327064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	521829	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	446935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	178340	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	173045	48.633	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.260%
35) Dibromofluoromethane	8.177	113	156170	48.326	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.660%
50) Toluene-d8	10.571	98	599797	49.641	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.280%
62) 4-Bromofluorobenzene	12.853	95	181836	45.437	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.880%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	15200	12.392	ug/l	96
52) Toluene	10.635	92	62959	7.299	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122222\
Data File : VN075947.D
Acq On : 22 Dec 2022 16:03
Operator : JC\MD
Sample : N6168-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1839

Quant Time: Dec 23 04:03:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

