

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122222\
 Data File : VN075950.D
 Acq On : 22 Dec 2022 17:24
 Operator : JC\MD
 Sample : N6151-05
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FS-01-(0-2)

Quant Time: Dec 23 05:24:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	299917	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	464800	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	391449	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	163308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	154949	47.489	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.980%
35) Dibromofluoromethane	8.171	113	137651	47.821	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.640%
50) Toluene-d8	10.571	98	536185	49.821	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.640%
62) 4-Bromofluorobenzene	12.853	95	155926	43.744	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.480%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.512	59	12577	27.242	ug/l #	72
52) Toluene	10.629	92	2309	0.301	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122222\
Data File : VN075950.D
Acq On : 22 Dec 2022 17:24
Operator : JC\MD
Sample : N6151-05
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FS-01-(0-2)

Quant Time: Dec 23 05:24:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

