

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122322\
 Data File : VN075955.D
 Acq On : 23 Dec 2022 10:27
 Operator : JC\MD
 Sample : VN1223WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1223WBL01

Quant Time: Dec 26 04:29:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	332577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	529605	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	457434	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	194942	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	176588	48.806	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.620%
35) Dibromofluoromethane	8.177	113	159476	48.624	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.240%
50) Toluene-d8	10.571	98	604924	49.330	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.660%
62) 4-Bromofluorobenzene	12.853	95	189580	46.677	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122322\
Data File : VN075955.D
Acq On : 23 Dec 2022 10:27
Operator : JC\MD
Sample : VN1223WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1223WBL01

Quant Time: Dec 26 04:29:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

