

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122322\
 Data File : VN075956.D
 Acq On : 23 Dec 2022 11:06
 Operator : JC\MD
 Sample : N6198-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-122222

Quant Time: Dec 26 04:29:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	306344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	490226	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	427598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	181851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	166624	49.996	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.000%
35) Dibromofluoromethane	8.171	113	149381	49.205	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.400%
50) Toluene-d8	10.571	98	566401	49.899	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.800%
62) 4-Bromofluorobenzene	12.853	95	178743	47.544	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.080%
Target Compounds						
						Qvalue
16) Acetone	4.448	43	19240	16.747	ug/l #	84
20) Methylene Chloride	5.289	84	3767	0.593	ug/l	87
25) 2-Butanone	7.500	43	3095	1.924	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122322\
Data File : VN075956.D
Acq On : 23 Dec 2022 11:06
Operator : JC\MD
Sample : N6198-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-01-122222

Quant Time: Dec 26 04:29:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

