

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122322\
 Data File : VN075967.D
 Acq On : 23 Dec 2022 15:33
 Operator : JC\MD
 Sample : PB149839ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149839ZHE#02

Quant Time: Dec 26 04:34:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	319676	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	529407	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	481378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	197152	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	173888	50.000	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.000%	
35) Dibromofluoromethane	8.177	113	155326	47.377	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.760%	
50) Toluene-d8	10.571	98	620370	50.609	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.220%	
62) 4-Bromofluorobenzene	12.853	95	197617	48.674	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.340%	
Target Compounds						
						Qvalue
16) Acetone	4.447	43	58777	49.028	ug/l	95
20) Methylene Chloride	5.283	84	12565	3.320	ug/l #	84
25) 2-Butanone	7.500	43	18660	11.116	ug/l	94
37) Ethyl Acetate	7.577	43	33548	9.590	ug/l	99
43) Isopropyl Acetate	8.694	43	305581	52.954	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122322\
Data File : VN075967.D
Acq On : 23 Dec 2022 15:33
Operator : JC\MD
Sample : PB149839ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149839ZHE#02

Quant Time: Dec 26 04:34:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

