

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122322\
 Data File : VN075972.D
 Acq On : 23 Dec 2022 17:34
 Operator : JC\MD
 Sample : N6206-01 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Dec 26 04:35:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	318123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	511232	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	430112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	181625	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	169337	48.929	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.860%
35) Dibromofluoromethane	8.177	113	154039	48.654	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.300%
50) Toluene-d8	10.571	98	585289	49.445	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.880%
62) 4-Bromofluorobenzene	12.853	95	176596	45.043	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.080%
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.606	83	1634	0.322	ug/l #	92
67) Ethyl Benzene	11.970	91	3821	0.256	ug/l #	84
73) Isopropylbenzene	12.700	105	5751	0.438	ug/l	91
78) n-propylbenzene	13.041	91	12765	0.882	ug/l	93
85) sec-Butylbenzene	13.617	105	4069	0.304	ug/l	86
89) n-Butylbenzene	14.064	91	2948	0.327	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122322\
Data File : VN075972.D
Acq On : 23 Dec 2022 17:34
Operator : JC\MD
Sample : N6206-01 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Dec 26 04:35:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

