

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122322\
 Data File : VN075958.D
 Acq On : 23 Dec 2022 11:54
 Operator : JC\MD
 Sample : N6198-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-122222

Quant Time: Dec 26 04:30:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	329654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	524501	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	466706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	195413	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	174949	48.782	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.560%
35) Dibromofluoromethane	8.177	113	159125	48.989	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.980%
50) Toluene-d8	10.571	98	617447	50.842	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.680%
62) 4-Bromofluorobenzene	12.853	95	191331	47.567	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.140%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122322\
Data File : VN075958.D
Acq On : 23 Dec 2022 11:54
Operator : JC\MD
Sample : N6198-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-122222

Quant Time: Dec 26 04:30:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.236 min Scan# 1050
Delta R.T. 0.006 min
Lab File: VN075958.D
Acq: 23 Dec 2022 11:54




Instrument :
MSVOA_N
ClientSampleId :
TB-01-122222

Tgt Ion:168 Resp: 329654
Ion Ratio Lower Upper
168 100
99 54.1 44.2 66.4



#33
1,2-Dichloroethane-d4
Concen: 48.782 ug/l
RT: 8.583 min Scan# 1110
Delta R.T. 0.000 min
Lab File: VN075958.D
Acq: 23 Dec 2022 11:54

Tgt Ion: 65 Resp: 174949
Ion Ratio Lower Upper
65 100
67 53.2 0.0 103.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1199

Delta R.T. 0.000 min

Lab File: VN075958.D

Acq: 23 Dec 2022 11:54

Instrument :

MSVOA_N

ClientSampleId :

TB-01-122222




Tgt Ion:114 Resp: 524501

Ion Ratio Lower Upper

114 100

63 17.4 0.0 34.4

88 14.6 0.0 28.8



#35

Dibromofluoromethane

Concen: 48.989 ug/l

RT: 8.177 min Scan# 1041

Delta R.T. 0.006 min

Lab File: VN075958.D

Acq: 23 Dec 2022 11:54




Tgt Ion:113 Resp: 159125


Ion Ratio Lower Upper

113 100

111 102.6 83.4 125.2

192 20.4 16.1 24.1









#50
Toluene-d8
Concen: 50.842 ug/l
RT: 10.571 min Scan# 1448
Delta R.T. 0.000 min
Lab File: VN075958.D
Acq: 23 Dec 2022 11:54




Instrument :
MSVOA_N
ClientSampleId :
TB-01-122222


Tgt Ion: 98 Resp: 617447
Ion Ratio Lower Upper
98 100
100 65.5 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 47.567 ug/l
RT: 12.853 min Scan# 1836
Delta R.T. 0.000 min
Lab File: VN075958.D
Acq: 23 Dec 2022 11:54

Tgt Ion: 95 Resp: 191331
Ion Ratio Lower Upper
95 100
174 86.1 0.0 171.6
176 82.7 0.0 169.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN075958.D
Acq: 23 Dec 2022 11:54

Instrument :
MSVOA_N
ClientSampleId :
TB-01-122222

Tgt Ion:117 Resp: 466706
Ion Ratio Lower Upper
117 100
82 52.4 41.3 61.9
119 32.1 24.6 37.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.800 min Scan# 1997
Delta R.T. 0.006 min
Lab File: VN075958.D
Acq: 23 Dec 2022 11:54

Tgt Ion:152 Resp: 195413
Ion Ratio Lower Upper
152 100
115 57.2 29.1 87.4
150 156.7 0.0 347.0

