

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122324\
 Data File : VN085287.D
 Acq On : 23 Dec 2024 12:02
 Operator : JC\MD
 Sample : VN1223MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1223MBL01

Quant Time: Dec 24 04:06:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 19 06:31:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	141668	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	254442	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	222830	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	88305	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	118343	47.793	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.580%
35) Dibromofluoromethane	8.159	113	90502	50.592	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.180%
50) Toluene-d8	10.565	98	313332	47.923	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.840%
62) 4-Bromofluorobenzene	12.841	95	101567	43.492	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	86.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122324\
Data File : VN085287.D
Acq On : 23 Dec 2024 12:02
Operator : JC\MD
Sample : VN1223MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1223MBL01

Quant Time: Dec 24 04:06:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 19 06:31:00 2024
Response via : Initial Calibration

