

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122324\
 Data File : VN085288.D
 Acq On : 23 Dec 2024 12:26
 Operator : JC\MD
 Sample : VN1223WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1223WBL01

Quant Time: Dec 24 04:06:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 19 06:31:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	174137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	324722	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	277455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	116189	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	149659	49.171	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.340%
35) Dibromofluoromethane	8.159	113	112526	49.289	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.580%
50) Toluene-d8	10.559	98	396960	47.573	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.140%
62) 4-Bromofluorobenzene	12.841	95	128030	42.958	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	85.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122324\
Data File : VN085288.D
Acq On : 23 Dec 2024 12:26
Operator : JC\MD
Sample : VN1223WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1223WBL01

Quant Time: Dec 24 04:06:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 19 06:31:00 2024
Response via : Initial Calibration

