

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122324\
 Data File : VN085292.D
 Acq On : 23 Dec 2024 14:26
 Operator : JC\MD
 Sample : PB165799TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB165799TB

Quant Time: Dec 24 04:09:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 19 06:31:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	149539	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	266221	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	227841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	93103	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	121211	46.375	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.760%		
35) Dibromofluoromethane	8.165	113	91202	48.728	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.460%		
50) Toluene-d8	10.565	98	325797	47.625	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.240%		
62) 4-Bromofluorobenzene	12.847	95	101787	41.658	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	83.320%		
Target Compounds						
16) Acetone	4.430	43	8252	9.147	ug/l #	88
29) Tetrahydrofuran	7.836	42	15383	17.272	ug/l	97
30) Chloroform	7.959	83	18414	4.721	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122324\
Data File : VN085292.D
Acq On : 23 Dec 2024 14:26
Operator : JC\MD
Sample : PB165799TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB165799TB

Quant Time: Dec 24 04:09:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 19 06:31:00 2024
Response via : Initial Calibration

