

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122324\
 Data File : VN085296.D
 Acq On : 23 Dec 2024 16:01
 Operator : JC\MD
 Sample : P5371-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-C11-1-M

Quant Time: Dec 24 04:12:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 19 06:31:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	190171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	333838	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	292239	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	135857	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	133303	40.104	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	80.200%	
35) Dibromofluoromethane	8.159	113	71661	30.532	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	61.060%#	
50) Toluene-d8	10.559	98	405728	47.296	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	94.600%	
62) 4-Bromofluorobenzene	12.841	95	150328	49.063	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	98.120%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	24536	21.385	ug/l #	88
27) cis-1,2-Dichloroethene	7.482	96	19158	6.876	ug/l	85
30) Chloroform	7.965	83	17662	3.560	ug/l	95
40) Benzene	8.600	78	201702	19.839	ug/l	97
44) Trichloroethene	9.359	130	2988	1.320	ug/l	65
52) Toluene	10.623	92	130890	21.411	ug/l	99
67) Ethyl Benzene	11.959	91	1676496	148.895	ug/l	98
68) m/p-Xylenes	12.070	106	151797	36.046	ug/l	96
69) o-Xylene	12.394	106	152990	38.651	ug/l	96
73) Isopropylbenzene	12.688	105	134451	13.719	ug/l	99
78) n-propylbenzene	13.029	91	70098	5.832	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	101910	12.467	ug/l	100
84) 1,2,4-Trimethylbenzene	13.476	105	381223	45.624	ug/l	98
86) p-Isopropyltoluene	13.723	119	19240	3.924	ug/l #	70
95) Naphthalene	15.635	128	13240403	1772.146	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122324\
Data File : VN085296.D
Acq On : 23 Dec 2024 16:01
Operator : JC\MD
Sample : P5371-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPLP-C11-1-M

Quant Time: Dec 24 04:12:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 19 06:31:00 2024
Response via : Initial Calibration

