

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122324\
 Data File : VN085303.D
 Acq On : 23 Dec 2024 18:47
 Operator : JC\MD
 Sample : P5371-01DL 4X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-C10-1-MDL

Quant Time: Dec 24 04:16:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 19 06:31:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	129978	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	242088	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	214104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	97477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	99852	43.953	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	87.900%		
35) Dibromofluoromethane	8.165	113	76848	45.152	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.300%		
50) Toluene-d8	10.565	98	301887	48.529	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.060%		
62) 4-Bromofluorobenzene	12.847	95	110184	49.590	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	99.180%		
Target Compounds					Qvalue	
16) Acetone	4.436	43	4406	5.619	ug/l	93
27) cis-1,2-Dichloroethene	7.483	96	4216	2.214	ug/l	85
30) Chloroform	7.965	83	5016	1.479	ug/l	96
40) Benzene	8.600	78	32609	4.423	ug/l	99
52) Toluene	10.624	92	56259	12.690	ug/l	99
67) Ethyl Benzene	11.959	91	345174	41.843	ug/l	97
68) m/p-Xylenes	12.065	106	51205	16.597	ug/l	98
69) o-Xylene	12.394	106	37622	12.973	ug/l	95
73) Isopropylbenzene	12.694	105	17631	2.507	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	14006	2.388	ug/l	86
84) 1,2,4-Trimethylbenzene	13.476	105	52938	8.830	ug/l	99
86) p-Isopropyltoluene	13.723	119	3917	2.508	ug/l #	71
95) Naphthalene	15.635	128	4435118	827.339	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122324\
Data File : VN085303.D
Acq On : 23 Dec 2024 18:47
Operator : JC\MD
Sample : P5371-01DL 4X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPLP-C10-1-MDL

Quant Time: Dec 24 04:16:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 19 06:31:00 2024
Response via : Initial Calibration

